

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q2106
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2106 **SAS No.:** Q2106

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168139BS							
Cadmium	ug/L	100	100		100	80 - 120	P
Copper	ug/L	150	159		106	80 - 120	P
Lead	ug/L	500	491		98	80 - 120	P
Nickel	ug/L	250	255		102	80 - 120	P
Zinc	ug/L	100	106		106	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q2106

Contract: ENTA05

Lab Code: CHEM

Case No.: Q2106

SAS No.: Q2106

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168168BS Mercury	ug/L	4.0	3.57		89	80 - 120	CV