

IDENTIFICATION SUMMARY  
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB168144BS

Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

Lab Sample ID: PB168144BS Date(s) Analyzed: 05/27/2025 05/27/2025

Instrument ID (1): ECD\_P Instrument ID (2): ECD\_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP072366.D

| ANALYTE      | COL | RT    | RT WINDOW |       | CONCENTRATION | MEAN<br>CONCENTRATION | %RPD  |
|--------------|-----|-------|-----------|-------|---------------|-----------------------|-------|
|              |     |       | FROM      | TO    |               |                       |       |
| Aroclor-1016 | 1   | 5.653 | 5.603     | 5.703 | 3.30          | 3.50                  | 5.56  |
| COLUMN 1     | 2   | 5.674 | 5.624     | 5.724 | 3.57          |                       |       |
|              | 3   | 5.736 | 5.686     | 5.786 | 3.46          |                       |       |
|              | 4   | 5.833 | 5.783     | 5.883 | 3.58          |                       |       |
|              | 5   | 6.126 | 6.076     | 6.176 | 3.43          |                       |       |
|              | 1   | 4.878 | 4.828     | 4.928 | 3.62          |                       |       |
| COLUMN 2     | 2   | 4.896 | 4.846     | 4.946 | 3.70          | 3.70                  |       |
|              | 3   | 5.073 | 5.023     | 5.123 | 3.75          |                       |       |
|              | 4   | 5.115 | 5.065     | 5.165 | 3.77          |                       |       |
|              | 5   | 5.329 | 5.279     | 5.379 | 3.68          |                       |       |
|              |     |       |           |       |               |                       |       |
| Aroclor-1260 | 1   | 7.243 | 7.193     | 7.293 | 3.69          | 3.50                  | 10.81 |
| COLUMN 1     | 2   | 7.497 | 7.447     | 7.547 | 3.68          |                       |       |
|              | 3   | 7.854 | 7.804     | 7.904 | 3.23          |                       |       |
|              | 4   | 8.078 | 8.028     | 8.128 | 3.52          |                       |       |
|              | 5   | 8.397 | 8.347     | 8.447 | 3.45          |                       |       |
|              | 1   | 6.361 | 6.311     | 6.411 | 4.05          |                       |       |
| COLUMN 2     | 2   | 6.549 | 6.499     | 6.599 | 3.87          | 3.90                  |       |
|              | 3   | 6.702 | 6.652     | 6.752 | 4.14          |                       |       |
|              | 4   | 7.172 | 7.122     | 7.222 | 3.69          |                       |       |
|              | 5   | 7.413 | 7.363     | 7.463 | 3.61          |                       |       |
|              |     |       |           |       |               |                       |       |

IDENTIFICATION SUMMARY  
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB168144BSD

Contract: ENTA05

Lab Code: CHEM

Case No.: Q2106

SAS No.: Q2106

SDG NO.: Q2106

Lab Sample ID: PB168144BSD

Date(s) Analyzed: 05/27/2025

05/27/2025

Instrument ID (1): ECD\_P

Instrument ID (2): ECD\_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP072367.D

| ANALYTE      | COL | RT    | RT WINDOW |       | CONCENTRATION | MEAN<br>CONCENTRATION | %RPD  |
|--------------|-----|-------|-----------|-------|---------------|-----------------------|-------|
|              |     |       | FROM      | TO    |               |                       |       |
| Aroclor-1016 | 1   | 5.654 | 5.604     | 5.704 | 3.32          | 3.50                  | 8.22  |
| COLUMN 1     | 2   | 5.675 | 5.625     | 5.725 | 3.56          |                       |       |
|              | 3   | 5.737 | 5.687     | 5.787 | 3.50          |                       |       |
|              | 4   | 5.834 | 5.784     | 5.884 | 3.57          |                       |       |
|              | 5   | 6.126 | 6.076     | 6.176 | 3.42          |                       |       |
|              | 1   | 4.879 | 4.829     | 4.929 | 3.81          |                       |       |
| COLUMN 2     | 2   | 4.897 | 4.847     | 4.947 | 3.79          | 3.80                  |       |
|              | 3   | 5.074 | 5.024     | 5.124 | 3.85          |                       |       |
|              | 4   | 5.115 | 5.065     | 5.165 | 3.85          |                       |       |
|              | 5   | 5.33  | 5.28      | 5.38  | 3.86          |                       |       |
|              |     |       |           |       |               |                       |       |
| Aroclor-1260 | 1   | 7.244 | 7.194     | 7.294 | 3.70          | 3.50                  | 10.81 |
| COLUMN 1     | 2   | 7.497 | 7.447     | 7.547 | 3.69          |                       |       |
|              | 3   | 7.855 | 7.805     | 7.905 | 3.17          |                       |       |
|              | 4   | 8.079 | 8.029     | 8.129 | 3.44          |                       |       |
|              | 5   | 8.398 | 8.348     | 8.448 | 3.43          |                       |       |
|              | 1   | 6.362 | 6.312     | 6.412 | 4.14          |                       |       |
| COLUMN 2     | 2   | 6.55  | 6.5       | 6.6   | 3.96          | 3.90                  |       |
|              | 3   | 6.702 | 6.652     | 6.752 | 4.20          |                       |       |
|              | 4   | 7.173 | 7.123     | 7.223 | 3.78          |                       |       |
|              | 5   | 7.415 | 7.365     | 7.465 | 3.57          |                       |       |
|              |     |       |           |       |               |                       |       |