

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

Continuing Calib Date: 05/23/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 12:26 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1232-1	(1)	4.86	4.86	4.76	4.96	0.00
Aroclor-1232-2	(2)	5.39	5.39	5.29	5.49	0.00
Aroclor-1232-3	(3)	5.67	5.68	5.58	5.78	0.01
Aroclor-1232-4	(4)	5.83	5.84	5.74	5.94	0.01
Aroclor-1232-5	(5)	5.92	5.93	5.83	6.03	0.01
Tetrachloro-m-xylene		4.50	4.50	4.40	4.60	0.00
Decachlorobiphenyl		10.21	10.21	10.11	10.31	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

Continuing Calib Date: 05/23/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 12:26 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1232-1	(1)	4.17	4.17	4.07	4.27	0.00
Aroclor-1232-2	(2)	4.90	4.90	4.80	5.00	0.00
Aroclor-1232-3	(3)	5.07	5.08	4.98	5.18	0.01
Aroclor-1232-4	(4)	5.16	5.16	5.06	5.26	0.00
Aroclor-1232-5	(5)	5.33	5.33	5.23	5.43	0.00
Tetrachloro-m-xylene		3.80	3.80	3.70	3.90	0.00
Decachlorobiphenyl		8.82	8.82	8.72	8.92	0.00