

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

Continuing Calib Date: 05/23/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 17:38 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1232-1 (1)   | 4.87       | 4.86      | 4.76      | 4.96  | 0.00       |
| Aroclor-1232-2 (2)   | 5.39       | 5.39      | 5.29      | 5.49  | 0.00       |
| Aroclor-1232-3 (3)   | 5.68       | 5.68      | 5.58      | 5.78  | 0.00       |
| Aroclor-1232-4 (4)   | 5.84       | 5.84      | 5.74      | 5.94  | 0.00       |
| Aroclor-1232-5 (5)   | 5.93       | 5.93      | 5.83      | 6.03  | 0.00       |
| Tetrachloro-m-xylene | 4.50       | 4.50      | 4.40      | 4.60  | 0.00       |
| Decachlorobiphenyl   | 10.21      | 10.21     | 10.11     | 10.31 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

Continuing Calib Date: 05/23/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 17:38 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1232-1       | (1) | 4.17       | 4.17      | 4.07      | 4.27 | 0.00       |
| Aroclor-1232-2       | (2) | 4.90       | 4.90      | 4.80      | 5.00 | 0.00       |
| Aroclor-1232-3       | (3) | 5.08       | 5.08      | 4.98      | 5.18 | 0.01       |
| Aroclor-1232-4       | (4) | 5.16       | 5.16      | 5.06      | 5.26 | 0.00       |
| Aroclor-1232-5       | (5) | 5.33       | 5.33      | 5.23      | 5.43 | 0.00       |
| Tetrachloro-m-xylene |     | 3.80       | 3.80      | 3.70      | 3.90 | 0.00       |
| Decachlorobiphenyl   |     | 8.82       | 8.82      | 8.72      | 8.92 | 0.00       |