

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/23/2025

Lab Sample No.: AR1660CCC500 Data File : PP072322.D Time Analyzed: 17:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.656	5.552	5.752	430.030	500.000	-14.0
Aroclor-1016-2	5.677	5.574	5.774	465.180	500.000	-7.0
Aroclor-1016-3	5.739	5.636	5.836	448.080	500.000	-10.4
Aroclor-1016-4	5.837	5.734	5.934	456.730	500.000	-8.7
Aroclor-1016-5	6.130	6.027	6.227	456.390	500.000	-8.7
Aroclor-1260-1	7.248	7.147	7.347	430.370	500.000	-13.9
Aroclor-1260-2	7.501	7.401	7.601	417.130	500.000	-16.6
Aroclor-1260-3	7.861	7.760	7.960	425.650	500.000	-14.9
Aroclor-1260-4	8.084	7.985	8.185	414.730	500.000	-17.1
Aroclor-1260-5	8.404	8.303	8.503	440.150	500.000	-12.0
Decachlorobiphenyl	10.211	10.113	10.313	44.670	50.000	-10.7
Tetrachloro-m-xylene	4.503	4.396	4.596	49.930	50.000	-0.1

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/23/2025

Lab Sample No.: AR1660CCC500 Data File : PP072322.D Time Analyzed: 17:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.880	4.777	4.977	518.370	500.000	3.7
Aroclor-1016-2	4.898	4.795	4.995	525.830	500.000	5.2
Aroclor-1016-3	5.075	4.973	5.173	537.100	500.000	7.4
Aroclor-1016-4	5.117	5.015	5.215	531.330	500.000	6.3
Aroclor-1016-5	5.331	5.229	5.429	566.760	500.000	13.4
Aroclor-1260-1	6.364	6.264	6.464	497.790	500.000	-0.4
Aroclor-1260-2	6.553	6.453	6.653	472.800	500.000	-5.4
Aroclor-1260-3	6.705	6.606	6.806	498.670	500.000	-0.3
Aroclor-1260-4	7.176	7.077	7.277	484.670	500.000	-3.1
Aroclor-1260-5	7.418	7.319	7.519	475.080	500.000	-5.0
Decachlorobiphenyl	8.820	8.723	8.923	50.170	50.000	0.3
Tetrachloro-m-xylene	3.797	3.691	3.891	52.470	50.000	4.9