

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/23/2025

Lab Sample No.: AR1660CCC500 Data File : PP072334.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.656	5.552	5.752	450.700	500.000	-9.9
Aroclor-1016-2	5.677	5.574	5.774	480.410	500.000	-3.9
Aroclor-1016-3	5.739	5.636	5.836	470.590	500.000	-5.9
Aroclor-1016-4	5.837	5.734	5.934	487.630	500.000	-2.5
Aroclor-1016-5	6.129	6.027	6.227	487.540	500.000	-2.5
Aroclor-1260-1	7.248	7.147	7.347	462.460	500.000	-7.5
Aroclor-1260-2	7.501	7.401	7.601	444.480	500.000	-11.1
Aroclor-1260-3	7.860	7.760	7.960	459.560	500.000	-8.1
Aroclor-1260-4	8.084	7.985	8.185	466.420	500.000	-6.7
Aroclor-1260-5	8.402	8.303	8.503	462.990	500.000	-7.4
Decachlorobiphenyl	10.209	10.113	10.313	47.720	50.000	-4.6
Tetrachloro-m-xylene	4.503	4.396	4.596	49.180	50.000	-1.6

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/23/2025

Lab Sample No.: AR1660CCC500 Data File : PP072334.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.879	4.777	4.977	497.610	500.000	-0.5
Aroclor-1016-2	4.897	4.795	4.995	503.890	500.000	0.8
Aroclor-1016-3	5.074	4.973	5.173	513.730	500.000	2.7
Aroclor-1016-4	5.116	5.015	5.215	512.180	500.000	2.4
Aroclor-1016-5	5.330	5.229	5.429	532.580	500.000	6.5
Aroclor-1260-1	6.363	6.264	6.464	519.730	500.000	3.9
Aroclor-1260-2	6.551	6.453	6.653	492.650	500.000	-1.5
Aroclor-1260-3	6.704	6.606	6.806	530.590	500.000	6.1
Aroclor-1260-4	7.175	7.077	7.277	530.330	500.000	6.1
Aroclor-1260-5	7.416	7.319	7.519	511.370	500.000	2.3
Decachlorobiphenyl	8.818	8.723	8.923	52.840	50.000	5.7
Tetrachloro-m-xylene	3.797	3.691	3.891	50.590	50.000	1.2