

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/27/2025

Lab Sample No.: AR1660CCC500 Data File : PP072360.D Time Analyzed: 08:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.651	5.552	5.752	437.760	500.000	-12.4
Aroclor-1016-2	5.672	5.574	5.774	470.780	500.000	-5.8
Aroclor-1016-3	5.734	5.636	5.836	458.890	500.000	-8.2
Aroclor-1016-4	5.832	5.734	5.934	476.120	500.000	-4.8
Aroclor-1016-5	6.124	6.027	6.227	466.060	500.000	-6.8
Aroclor-1260-1	7.242	7.147	7.347	455.490	500.000	-8.9
Aroclor-1260-2	7.495	7.401	7.601	458.540	500.000	-8.3
Aroclor-1260-3	7.853	7.760	7.960	475.440	500.000	-4.9
Aroclor-1260-4	8.077	7.985	8.185	483.230	500.000	-3.4
Aroclor-1260-5	8.396	8.303	8.503	496.300	500.000	-0.7
Decachlorobiphenyl	10.198	10.113	10.313	49.190	50.000	-1.6
Tetrachloro-m-xylene	4.499	4.396	4.596	49.100	50.000	-1.8

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/27/2025

Lab Sample No.: AR1660CCC500 Data File : PP072360.D Time Analyzed: 08:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.877	4.777	4.977	490.520	500.000	-1.9
Aroclor-1016-2	4.894	4.795	4.995	490.060	500.000	-2.0
Aroclor-1016-3	5.071	4.973	5.173	498.050	500.000	-0.4
Aroclor-1016-4	5.113	5.015	5.215	507.690	500.000	1.5
Aroclor-1016-5	5.327	5.229	5.429	499.330	500.000	-0.1
Aroclor-1260-1	6.359	6.264	6.464	496.680	500.000	-0.7
Aroclor-1260-2	6.548	6.453	6.653	474.140	500.000	-5.2
Aroclor-1260-3	6.700	6.606	6.806	501.950	500.000	0.4
Aroclor-1260-4	7.170	7.077	7.277	502.570	500.000	0.5
Aroclor-1260-5	7.412	7.319	7.519	498.600	500.000	-0.3
Decachlorobiphenyl	8.813	8.723	8.923	49.410	50.000	-1.2
Tetrachloro-m-xylene	3.795	3.691	3.891	49.920	50.000	-0.2