

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 05/27/2025

Lab Sample No.: AR1660CCC500 Data File : PP072375.D Time Analyzed: 15:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.655	5.552	5.752	428.750	500.000	-14.3
Aroclor-1016-2	5.676	5.574	5.774	460.120	500.000	-8.0
Aroclor-1016-3	5.738	5.636	5.836	453.000	500.000	-9.4
Aroclor-1016-4	5.836	5.734	5.934	471.630	500.000	-5.7
Aroclor-1016-5	6.128	6.027	6.227	463.270	500.000	-7.3
Aroclor-1260-1	7.246	7.147	7.347	456.340	500.000	-8.7
Aroclor-1260-2	7.499	7.401	7.601	448.890	500.000	-10.2
Aroclor-1260-3	7.857	7.760	7.960	473.370	500.000	-5.3
Aroclor-1260-4	8.081	7.985	8.185	467.550	500.000	-6.5
Aroclor-1260-5	8.399	8.303	8.503	477.050	500.000	-4.6
Decachlorobiphenyl	10.203	10.113	10.313	48.040	50.000	-3.9
Tetrachloro-m-xylene	4.503	4.396	4.596	47.500	50.000	-5.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 05/27/2025

Lab Sample No.: AR1660CCC500 Data File : PP072375.D Time Analyzed: 15:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.878	4.777	4.977	476.580	500.000	-4.7
Aroclor-1016-2	4.896	4.795	4.995	480.550	500.000	-3.9
Aroclor-1016-3	5.073	4.973	5.173	489.590	500.000	-2.1
Aroclor-1016-4	5.115	5.015	5.215	494.050	500.000	-1.2
Aroclor-1016-5	5.329	5.229	5.429	499.640	500.000	-0.1
Aroclor-1260-1	6.362	6.264	6.464	504.400	500.000	0.9
Aroclor-1260-2	6.550	6.453	6.653	478.560	500.000	-4.3
Aroclor-1260-3	6.703	6.606	6.806	508.440	500.000	1.7
Aroclor-1260-4	7.173	7.077	7.277	505.070	500.000	1.0
Aroclor-1260-5	7.415	7.319	7.519	504.470	500.000	0.9
Decachlorobiphenyl	8.815	8.723	8.923	52.670	50.000	5.3
Tetrachloro-m-xylene	3.797	3.691	3.891	49.200	50.000	-1.6