

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |               |                           |                   |
|-----------------|---------------|---------------------------|-------------------|
| <b>SDG No.:</b> | <u>Q2106</u>  | <b>Analytical Method:</b> | <u>8082A</u>      |
| <b>Client:</b>  | <u>ENTACT</u> | <b>Datafile :</b>         | <u>PP072367.D</u> |

| Lab Sample ID             | Parameter | Spike | Result | Units | Rec | RPD | Qual | RPD  | Limits  |           | RPD     |
|---------------------------|-----------|-------|--------|-------|-----|-----|------|------|---------|-----------|---------|
|                           |           |       |        |       |     |     |      | Qual | Low     | High      |         |
| PB168144BSD<br>(Column 1) | AR1016    | 5     | 3.50   | ug/L  | 70  | 0   |      |      | 40 (61) | 140 (112) | 20 (20) |
|                           | AR1260    | 5     | 3.50   | ug/L  | 70  | 0   |      |      | 40 (66) | 140 (113) | 20 (20) |
| PB168144BSD<br>(Column 2) | AR1016    | 5     | 3.80   | ug/L  | 76  | 3   |      |      | 40 (61) | 140 (112) | 20 (20) |
|                           | AR1260    | 5     | 3.90   | ug/L  | 78  | 0   |      |      | 40 (66) | 140 (113) | 20 (20) |