

Surrogate Summary

SW-846

SDG No.: Q2106

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168143BL	PB168143BL	2-Fluorophenol	150	125	84		15 (10)	110 (139)
		Phenol-d6	150	126	84		15 (10)	110 (134)
		Nitrobenzene-d5	100	78.4	78		30 (49)	130 (133)
		2-Fluorobiphenyl	100	73.4	73		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	133	89		15 (44)	110 (137)
		Terphenyl-d14	100	69.5	70		30 (48)	130 (125)
PB168143BS	PB168143BS	2-Fluorophenol	150	120	80		15 (10)	110 (139)
		Phenol-d6	150	124	83		15 (10)	110 (134)
		Nitrobenzene-d5	100	75.5	76		30 (49)	130 (133)
		2-Fluorobiphenyl	100	71.3	71		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	132	88		15 (44)	110 (137)
		Terphenyl-d14	100	75.5	75		30 (48)	130 (125)
PB168143BSD	PB168143BSD	2-Fluorophenol	150	119	79		15 (10)	110 (139)
		Phenol-d6	150	121	80		15 (10)	110 (134)
		Nitrobenzene-d5	100	73.8	74		30 (49)	130 (133)
		2-Fluorobiphenyl	100	70.7	71		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	127	84		15 (44)	110 (137)
		Terphenyl-d14	100	75.0	75		30 (48)	130 (125)
Q2106-01	TW-WTS-09	2-Fluorophenol	150	61.0	41		15 (10)	110 (139)
		Phenol-d6	150	34.5	23		15 (10)	110 (134)
		Nitrobenzene-d5	100	69.7	70		30 (49)	130 (133)
		2-Fluorobiphenyl	100	70.5	71		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	152	101		15 (44)	110 (137)
		Terphenyl-d14	100	81.0	81		30 (48)	130 (125)