

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/23/2025

Lab File ID: BP024788.D Time Analyzed: 11:40

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	85825	7.646	348651	10.42	223501	14.28
UPPER LIMIT	171650	8.146	697302	10.916	447002	14.781
LOWER LIMIT	42912.5	7.146	174326	9.916	111751	13.781
EPA SAMPLE NO.						
01 TW-WTS-09	99275	7.65	401101	10.42	275482	14.28

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.