

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG No.: Q2106

Instrument ID: MSVOA_X Calibration Date/Time: 05/23/2025 09:24

Lab File ID: VX046331.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
1,2-Dichloroethane-d4	0.932	0.966		3.65	20
Dibromofluoromethane	0.360	0.384		6.67	20
Toluene-d8	1.246	1.307		4.9	20
4-Bromofluorobenzene	0.478	0.514		7.53	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.