

Report of Analysis

Client:	ENTACT		Date Collected:		
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:		
Client Sample ID:	VX0523WBSD01		SDG No.:	Q2106	
Lab Sample ID:	VX0523WBSD01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group4	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046337.D	1		05/23/25 12:10	VX052325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	20.8		0.16	1.00	ug/L
56-23-5	Carbon Tetrachloride	19.1		0.25	1.00	ug/L
67-66-3	Chloroform	21.0		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.4		0.20	1.00	ug/L
71-43-2	Benzene	20.4		0.15	1.00	ug/L
108-88-3	Toluene	20.7		0.14	1.00	ug/L
127-18-4	Tetrachloroethene	20.2		0.23	1.00	ug/L
100-41-4	Ethyl Benzene	20.7		0.13	1.00	ug/L
1330-20-7	Total Xylenes	62.1		0.36	3.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.2		70 (74) - 130 (125)	100%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		70 (75) - 130 (124)	102%	SPK: 50
2037-26-5	Toluene-d8	50.3		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.8		70 (77) - 130 (121)	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	88800	5.55			
540-36-3	1,4-Difluorobenzene	156000	6.757			
3114-55-4	Chlorobenzene-d5	139000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	64500	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products