

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2106 SAS No.: Q2106 SDG NO.: Q2106
 Lab File ID: VX046331.D Date Analyzed: 05/23/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:24
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	89943	5.54	154938	6.76	137123	10.05
UPPER LIMIT	179886	6.044	309876	7.257	274246	10.549
LOWER LIMIT	44971.5	5.044	77469	6.257	68561.5	9.549
EPA SAMPLE NO.						
TW-WTS-09	61490	5.54	124971	6.76	119162	10.05
VX0523WBL01	69169	5.55	136383	6.76	129888	10.05
VX0523WBS01	88801	5.54	154286	6.75	135355	10.05
VX0523WBSD01	88781	5.55	156018	6.76	138561	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.