

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q2114 SAS No.: Q2114 SDG No.: Q2114

Instrument ID: BNA_F Calibration Date/Time: 05/23/2025 11:12

Lab File ID: BF142527.D Init. Calib. Date(s): 05/20/2025 05/20/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:10 15:31

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.187	1.152		-2.9	
Benzaldehyde	0.859	0.837		-2.6	
Phenol-d6	1.429	1.392		-2.6	
bis(2-Chloroethyl)ether	1.157	1.112		-3.9	
2,2-oxybis(1-Chloropropane)	1.944	1.901		-2.2	
Acetophenone	0.443	0.427		-3.6	
n-Nitroso-di-n-propylamine	0.886	0.848	0.050	-4.3	
Nitrobenzene-d5	0.367	0.350		-4.6	
Hexachloroethane	0.507	0.478		-5.7	
Nitrobenzene	0.331	0.318		-3.9	
Isophorone	0.613	0.580		-5.4	
bis(2-Chloroethoxy)methane	0.383	0.365		-4.7	
Naphthalene	0.985	0.937		-4.9	
4-Chloroaniline	0.399	0.390		-2.3	
Hexachlorobutadiene	0.192	0.178		-7.3	20.0
Caprolactam	0.080	0.084		5.0	
2-Methylnaphthalene	0.620	0.586		-5.5	
Hexachlorocyclopentadiene	0.387	0.374	0.050	-3.4	
2-Fluorobiphenyl	1.490	1.386		-7.0	
1,1-Biphenyl	1.552	1.477		-4.8	
2-Chloronaphthalene	1.146	1.093		-4.6	
2-Nitroaniline	0.331	0.326		-1.5	
Dimethylphthalate	1.320	1.270		-3.8	
Acenaphthylene	1.936	1.847		-4.6	
2,6-Dinitrotoluene	0.285	0.275		-3.5	
3-Nitroaniline	0.311	0.314		1.0	
Acenaphthene	1.182	1.124		-4.9	20.0
Dibenzofuran	1.701	1.625		-4.5	
2,4-Dinitrotoluene	0.372	0.376		1.1	
Diethylphthalate	1.289	1.248		-3.2	
4-Chlorophenyl-phenylether	0.646	0.620		-4.0	
Fluorene	1.314	1.263		-3.9	
4-Nitroaniline	0.282	0.300		6.4	
n-Nitrosodiphenylamine	0.684	0.633		-7.5	20.0
2,4,6-Tribromophenol	0.222	0.216		-2.7	
4-Bromophenyl-phenylether	0.236	0.214		-9.3	
Hexachlorobenzene	0.262	0.242		-7.6	
Atrazine	0.182	0.183		0.5	
Phenanthrene	1.069	1.006		-5.9	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q2114 SAS No.: Q2114 SDG No.: Q2114

Instrument ID: BNA_F Calibration Date/Time: 05/23/2025 11:12

Lab File ID: BF142527.D Init. Calib. Date(s): 05/20/2025 05/20/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:10 15:31

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Anthracene	1.091	1.036		-5.0	
Carbazole	0.933	0.921		-1.3	
Di-n-butylphthalate	1.021	1.043		2.2	
Fluoranthene	0.999	0.980		-1.9	20.0
Pyrene	1.866	1.802		-3.4	
Terphenyl-d14	1.464	1.397		-4.6	
Butylbenzylphthalate	0.516	0.553		7.2	
3,3-Dichlorobenzidine	0.405	0.413		2.0	
Benzo(a)anthracene	1.336	1.324		-0.9	
Chrysene	1.200	1.134		-5.5	
Bis(2-ethylhexyl)phthalate	0.660	0.677		2.6	
Di-n-octyl phthalate	1.290	1.118		-13.3	20.0
Benzo(b)fluoranthene	1.184	1.171		-1.1	
Benzo(k)fluoranthene	1.109	0.983		-11.4	
Benzo(a)pyrene	1.116	1.065		-4.6	20.0
Indeno(1,2,3-cd)pyrene	1.501	1.373		-8.5	
Dibenzo(a,h)anthracene	1.218	1.112		-8.7	
Benzo(g,h,i)perylene	1.218	1.127		-7.5	
1,2,4,5-Tetrachlorobenzene	0.574	0.550		-4.2	
1,4-Dioxane	0.475	0.471		-0.8	20.0

All other compounds must meet a minimum RRF of 0.010.