

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01  
 Lab Code: CHEM Case No.: Q2114 SAS No.: Q2114 SDG No.: Q2114  
 Instrument ID: MSVOA\_X Calibration Date(s): 05/05/2025 05/05/2025  
 Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27  
 GC Column: DB-624UI ID: 0.18 (mm)

| LAB FILE ID:                   |        | RRF020 = VX046041.D |        | RRF050 = VX046042.D |        | RRF100 = VX046043.D |       |       |  |
|--------------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                                |        | RRF150 = VX046044.D |        | RRF005 = VX046046.D |        | RRF001 = VX046047.D |       |       |  |
| COMPOUND                       | RRF020 | RRF050              | RRF100 | RRF150              | RRF005 | RRF001              | RRF   | % RSD |  |
| Dichlorodifluoromethane        | 0.697  | 0.864               | 0.859  | 0.875               | 0.639  | 0.658               | 0.765 | 14.6  |  |
| Chloromethane                  | 0.727  | 0.775               | 0.787  | 0.791               | 0.679  | 0.694               | 0.742 | 6.6   |  |
| Vinyl Chloride                 | 0.660  | 0.710               | 0.727  | 0.755               | 0.619  | 0.673               | 0.691 | 7.2   |  |
| Bromomethane                   | 0.296  | 0.326               | 0.340  | 0.334               | 0.305  |                     | 0.320 | 5.8   |  |
| Chloroethane                   | 0.354  | 0.378               | 0.329  | 0.317               | 0.368  | 0.467               | 0.369 | 14.4  |  |
| Trichlorofluoromethane         | 1.035  | 1.068               | 0.983  | 0.985               | 0.990  | 1.064               | 1.021 | 3.9   |  |
| 1,1,2-Trichlorotrifluoroethane | 0.628  | 0.641               | 0.629  | 0.648               | 0.610  | 0.633               | 0.632 | 2.1   |  |
| 1,1-Dichloroethene             | 0.565  | 0.601               | 0.607  | 0.625               | 0.567  | 0.594               | 0.593 | 3.9   |  |
| Acetone                        | 0.361  | 0.362               | 0.361  | 0.370               | 0.408  | 0.380               | 0.374 | 4.9   |  |
| Carbon Disulfide               | 1.295  | 1.455               | 1.522  | 1.597               | 1.141  | 1.423               | 1.406 | 11.7  |  |
| Methyl tert-butyl Ether        | 2.044  | 2.160               | 2.172  | 2.239               | 1.908  | 1.949               | 2.079 | 6.4   |  |
| Methyl Acetate                 | 0.814  | 0.848               | 0.845  | 0.875               | 0.816  | 1.006               | 0.867 | 8.3   |  |
| Methylene Chloride             | 0.689  | 0.684               | 0.691  | 0.691               | 0.689  | 0.853               | 0.716 | 9.4   |  |
| trans-1,2-Dichloroethene       | 0.573  | 0.610               | 0.612  | 0.622               | 0.557  | 0.604               | 0.596 | 4.3   |  |
| 1,1-Dichloroethane             | 1.233  | 1.263               | 1.263  | 1.286               | 1.154  | 1.116               | 1.219 | 5.6   |  |
| Cyclohexane                    | 1.090  | 1.128               | 1.128  | 1.150               | 1.059  |                     | 1.111 | 3.3   |  |
| 2-Butanone                     | 0.540  | 0.555               | 0.558  | 0.569               | 0.539  | 0.495               | 0.543 | 4.8   |  |
| Carbon Tetrachloride           | 0.528  | 0.558               | 0.552  | 0.577               | 0.505  | 0.541               | 0.544 | 4.6   |  |
| cis-1,2-Dichloroethene         | 0.716  | 0.737               | 0.738  | 0.755               | 0.642  | 0.719               | 0.718 | 5.5   |  |
| Bromochloromethane             | 0.628  | 0.578               | 0.595  | 0.590               | 0.553  | 0.576               | 0.587 | 4.3   |  |
| Chloroform                     | 1.287  | 1.296               | 1.277  | 1.300               | 1.199  | 1.265               | 1.271 | 3     |  |
| 1,1,1-Trichloroethane          | 1.106  | 1.131               | 1.155  | 1.188               | 1.013  | 1.015               | 1.101 | 6.6   |  |
| Methylcyclohexane              | 0.596  | 0.641               | 0.627  | 0.658               | 0.587  | 0.627               | 0.623 | 4.3   |  |
| Benzene                        | 1.426  | 1.474               | 1.441  | 1.477               | 1.337  | 1.348               | 1.417 | 4.3   |  |
| 1,2-Dichloroethane             | 0.632  | 0.627               | 0.611  | 0.625               | 0.594  | 0.579               | 0.612 | 3.5   |  |
| Trichloroethene                | 0.344  | 0.355               | 0.345  | 0.362               | 0.315  | 0.324               | 0.341 | 5.3   |  |
| 1,2-Dichloropropane            | 0.356  | 0.371               | 0.368  | 0.378               | 0.324  | 0.317               | 0.352 | 7.4   |  |
| Bromodichloromethane           | 0.557  | 0.577               | 0.573  | 0.594               | 0.498  | 0.485               | 0.547 | 8.2   |  |
| 4-Methyl-2-Pentanone           | 0.620  | 0.634               | 0.630  | 0.631               | 0.555  | 0.561               | 0.605 | 6     |  |
| Toluene                        | 0.884  | 0.898               | 0.885  | 0.904               | 0.838  | 0.803               | 0.869 | 4.5   |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01  
Lab Code: CHEM Case No.: Q2114 SAS No.: Q2114 SDG No.: Q2114  
Instrument ID: MSVOA\_X Calibration Date(s): 05/05/2025 05/05/2025  
Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27  
GC Column: DB-624UI ID: 0.18 (mm)

| LAB FILE ID:                |        | RRF020 = VX046041.D |        | RRF050 = VX046042.D |        | RRF100 = VX046043.D |       |       |  |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                             |        | RRF150 = VX046044.D |        | RRF005 = VX046046.D |        | RRF001 = VX046047.D |       |       |  |
| COMPOUND                    | RRF020 | RRF050              | RRF100 | RRF150              | RRF005 | RRF001              | RRF   | % RSD |  |
| t-1,3-Dichloropropene       | 0.468  | 0.528               | 0.555  | 0.591               | 0.406  | 0.371               | 0.487 | 17.9  |  |
| cis-1,3-Dichloropropene     | 0.531  | 0.578               | 0.602  | 0.623               | 0.469  | 0.423               | 0.538 | 14.6  |  |
| 1,1,2-Trichloroethane       | 0.349  | 0.354               | 0.351  | 0.356               | 0.337  | 0.308               | 0.343 | 5.3   |  |
| 2-Hexanone                  | 0.466  | 0.473               | 0.477  | 0.473               | 0.414  | 0.385               | 0.448 | 8.7   |  |
| Dibromochloromethane        | 0.378  | 0.400               | 0.415  | 0.431               | 0.326  | 0.306               | 0.376 | 13.3  |  |
| 1,2-Dibromoethane           | 0.359  | 0.373               | 0.368  | 0.381               | 0.333  | 0.322               | 0.356 | 6.5   |  |
| Tetrachloroethene           | 0.390  | 0.375               | 0.345  | 0.344               | 0.323  | 0.347               | 0.354 | 6.8   |  |
| Chlorobenzene               | 1.093  | 1.098               | 1.085  | 1.114               | 1.046  | 1.131               | 1.094 | 2.7   |  |
| Ethyl Benzene               | 1.919  | 2.022               | 1.979  | 2.036               | 1.816  | 1.803               | 1.929 | 5.2   |  |
| m/p-Xylenes                 | 0.706  | 0.740               | 0.721  | 0.740               | 0.678  | 0.648               | 0.706 | 5.2   |  |
| o-Xylene                    | 0.688  | 0.727               | 0.706  | 0.726               | 0.639  | 0.642               | 0.688 | 5.7   |  |
| Styrene                     | 1.135  | 1.219               | 1.214  | 1.230               | 1.012  | 0.951               | 1.127 | 10.6  |  |
| Bromoform                   | 0.270  | 0.304               | 0.312  | 0.327               | 0.236  | 0.234               | 0.281 | 14.2  |  |
| Isopropylbenzene            | 3.843  | 4.130               | 3.876  | 4.156               | 3.562  | 3.789               | 3.893 | 5.7   |  |
| 1,1,2,2-Tetrachloroethane   | 1.315  | 1.338               | 1.284  | 1.345               | 1.350  | 1.552               | 1.364 | 7     |  |
| 1,2,4-Trimethylbenzene      | 3.274  | 3.522               | 3.335  | 3.444               | 3.034  | 3.150               | 3.293 | 5.5   |  |
| 1,3-Dichlorobenzene         | 1.633  | 1.701               | 1.656  | 1.730               | 1.558  | 1.619               | 1.649 | 3.7   |  |
| 1,4-Dichlorobenzene         | 1.629  | 1.693               | 1.639  | 1.722               | 1.606  | 1.817               | 1.684 | 4.6   |  |
| 1,2-Dichlorobenzene         | 1.613  | 1.696               | 1.634  | 1.702               | 1.577  | 1.710               | 1.655 | 3.3   |  |
| 1,2-Dibromo-3-Chloropropane | 0.299  | 0.322               | 0.329  | 0.356               | 0.248  | 0.259               | 0.302 | 13.9  |  |
| 1,2,4-Trichlorobenzene      | 0.861  | 0.981               | 1.035  | 1.123               | 0.842  | 0.862               | 0.951 | 12    |  |
| 1,2,3-Trichlorobenzene      | 0.921  | 1.019               | 1.051  | 1.107               | 0.846  | 0.941               | 0.981 | 9.7   |  |
| 1,2-Dichloroethane-d4       | 0.953  | 0.910               | 0.930  | 0.932               | 0.935  |                     | 0.932 | 1.6   |  |
| Dibromofluoromethane        | 0.359  | 0.355               | 0.364  | 0.368               | 0.354  |                     | 0.360 | 1.7   |  |
| Toluene-d8                  | 1.246  | 1.223               | 1.266  | 1.275               | 1.221  |                     | 1.246 | 2     |  |
| 4-Bromofluorobenzene        | 0.455  | 0.470               | 0.500  | 0.500               | 0.464  |                     | 0.478 | 4.4   |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.