

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q2115

Project: AEC-2025-0013- 91 Fulton, White Plains Bus Co

RunNo.: LB135932

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: COD	ICV	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: COD	CCV1	mg/L	48.931	50	98	95-105	05/28/2025
Sample ID: COD	CCV2	mg/L	50.962	50	102	95-105	05/28/2025