

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ATGG01

Lab Code: CHEM Case No.: Q2115 SAS No.: Q2115 SDG No.: Q2115

Instrument ID: MSVOA_N Calibration Date/Time: 05/23/2025 09:53

Lab File ID: VN086766.D Init. Calib. Date(s): 05/16/2025 05/16/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 16:31 18:32

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.433	1.390		-3	20
Toluene	0.891	0.879		-1.35	20
Ethyl Benzene	1.878	1.879		0.05	20
m/p-Xylenes	0.711	0.744		4.64	20
o-Xylene	0.707	0.732		3.54	20
1,2-Dichloroethane-d4	0.691	0.591		-14.47	20
Dibromofluoromethane	0.298	0.303		1.68	20
Toluene-d8	1.234	1.244		0.81	20
4-Bromofluorobenzene	0.436	0.447		2.52	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.