

Surrogate Summary

SDG No.: Q2115

Client: ATG-GREENVILLE AEC

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2115-01	OUTFALL-1	1,2-Dichloroethane-d4	50	44.8	90	74	125
		Dibromofluoromethane	50	50.7	101	75	124
		Toluene-d8	50	51.1	102	86	113
		4-Bromofluorobenzene	50	53.9	108	77	121
VN0523WBL01	VN0523WBL01	1,2-Dichloroethane-d4	50	42.6	85	74	125
		Dibromofluoromethane	50	50.3	101	75	124
		Toluene-d8	50	50.9	102	86	113
		4-Bromofluorobenzene	50	51.9	104	77	121
VN0523WBS01	VN0523WBS01	1,2-Dichloroethane-d4	50	41.9	84	74	125
		Dibromofluoromethane	50	49.6	99	75	124
		Toluene-d8	50	49.6	99	86	113
		4-Bromofluorobenzene	50	49.9	100	77	121
VN0523WBSD0	VN0523WBSD01	1,2-Dichloroethane-d4	50	44.8	90	74	125
		Dibromofluoromethane	50	50.2	100	75	124
		Toluene-d8	50	48.9	98	86	113
		4-Bromofluorobenzene	50	49.1	98	77	121