

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2116
Project:	AEC-2025-0013- 14 Fisher Lane, White Plains Bus Co	Sample ID:	Q2116-01
Client ID:	OUTFALL-2MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	65.2		18.5		50.0	1	93		05/28/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2116
Project:	AEC-2025-0013- 14 Fisher Lane, White Plains Bus Co	Sample ID:	Q2116-01
Client ID:	OUTFALL-2MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	64.2		18.5		50.0	1	91		05/28/2025