

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2116

Client: ATG-GREENVILLE AEC

Analytical Method: SW8260-Low

Datafile : VN086769.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0523WBS01	Benzene	20	17.5	ug/L	88			82	109	
	Toluene	20	18.2	ug/L	91			82	110	
	Ethyl Benzene	20	18.3	ug/L	92			83	109	
	m/p-Xylenes	40	39.0	ug/L	98			82	110	
	o-Xylene	20	19.1	ug/L	96			83	109	