

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ATGG01
 Lab Code: CHEM Case No.: Q2116 SAS No.: Q2116 SDG NO.: Q2116
 Lab File ID: VN086766.D Date Analyzed: 05/23/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:53
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	276686	8.22	466652	9.09	409394	11.87
UPPER LIMIT	553372	8.718	933304	9.594	818788	12.365
LOWER LIMIT	138343	7.718	233326	8.594	204697	11.365
EPA SAMPLE NO.						
OUTFALL-2	246001	8.22	452072	9.10	437261	11.87
VN0523WBL01	261056	8.22	476062	9.10	449746	11.87
VN0523WBS01	283024	8.22	485622	9.10	427228	11.87
VN0523WBSD01	237879	8.22	417988	9.09	365055	11.86

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.