

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ATGG01
 Lab Code: CHEM Case No.: Q2116 SAS No.: Q2116 SDG NO.: Q2116
 Lab File ID: VN086766.D Date Analyzed: 05/23/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:53
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	193032	13.788				
UPPER LIMIT	386064	14.288				
LOWER LIMIT	96516	13.288				
EPA SAMPLE NO.						
OUTFALL-2	192791	13.79				
VN0523WBL01	197068	13.79				
VN0523WBS01	194058	13.79				
VN0523WBSD01	171389	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.