# ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

### GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2118	
-----------------	--

MATRIX: /Water

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			$\checkmark$
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.			$\checkmark$
	The Initial Calibration met the requirements. The Continuous Calibration File ID VY022407.D met the requirements except for Bromomethane,Chloroethane and Vinyl Chloride are failing high but no positive hit in associate sample therefore no corrective action taken.			
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		$\checkmark$	
7.	Surrogate Recoveries Meet Criteria		$\checkmark$	
	If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
	The Surrogate recoveries met the acceptable criteria except for BP-VPB-182-GW- 500-502 [4-Bromofluorobenzene - 77%], BP-VPB-182-GW-500-502RE [4- Bromofluorobenzene - 76%] sample was reanalyzed to confirm the failure and reported.			
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria			$\checkmark$
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The Blank Spike for {VY0523SBS01} with File ID: VY022409.D met requirements for			
	all samples except for Bromomethane[173%] is failing high but no positive hit in associat	ie		
	sample therefore no corrective action taken.			
	The Blank Spike Duplicate for {VY0523SBSD01} with File ID: VY022410.D met			
	requirements for all samples except for Bromomethane[187%], Chloroethane[155%] are			

failing high but no positive hit in associate sample therefore no corrective action taken.

## ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

### GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

		NA	NO	YES
9.	Internal Standard Area/Retention Time Shift Meet Criteria			$\checkmark$
	Comments:			
10.	Analysis Holding Time Met			$\checkmark$
	If not met, list number of days exceeded for each sample:			

### ADDITIONAL COMMENTS:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The laboratory certifies that the all-electronic diskette deliverable exactly match the data Summary forms (i.e. Form Is)."

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the % RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is >20% for the Initial Calibration curve for SW-846 analysis.

**QA REVIEW** 

Date