

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2118

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified **DataFile:** BN037109.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168155BSD	1,4-Dioxane	0.4	0.30	ug/L	75	3			70	130	20