

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2118 SAS No.: Q2118 SDG NO.: Q2118

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/29/2025

Lab File ID: BN037128.D Time Analyzed: 10:07

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2387	7.604	6582	10.38	3682	14.26
UPPER LIMIT	4774	8.104	13164	10.883	7364	14.756
LOWER LIMIT	1193.5	7.104	3291	9.883	1841	13.756
EPA SAMPLE NO.						
01 BP-VPB-182-GW-540-542	2218	7.60	5955	10.38	3394	14.26

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.