

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2118 SAS No.: Q2118 SDG NO.: Q2118
Lab File ID: VY022424.D Date Analyzed: 05/27/2025
Instrument ID: MSVOA_Y Time Analyzed: 09:59
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	251462	7.71	421648	8.62	359387	11.42
UPPER LIMIT	502924	8.213	843296	9.116	718774	11.92
LOWER LIMIT	125731	7.213	210824	8.116	179694	10.92
EPA SAMPLE NO.						
BP-VPB-182-GW-500-502RE	322318	7.71	572594	8.62	436021	11.41
VY0527SBL01	351784	7.71	631221	8.62	510103	11.41
VY0527SBS01	230468	7.71	399143	8.62	340460	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.