

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2118 SAS No.: Q2118 SDG NO.: Q2118
Lab File ID: VX046331.D Date Analyzed: 05/23/2025
Instrument ID: MSVOA_X Time Analyzed: 09:24
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	66629	12.018				
UPPER LIMIT	133258	12.518				
LOWER LIMIT	33314.5	11.518				
EPA SAMPLE NO.						
BP-VPB-182-TB-20250520	55490	12.02				
BP-VPB-182-GW-420-422	49690	12.02				
BP-VPB-182-GW-450-452	49143	12.02				
BP-VPB-182-DUP-20250520	47836	12.02				
BP-VPB-182-GW-460-462	54701	12.02				
BP-VPB-182-GW-480-482	55196	12.02				
BP-VPB-182-GW-520-522	52504	12.02				
BP-VPB-182-GW-540-542	53639	12.02				
BP-VPB-182-GW-560-562	49123	12.02				
VX0523WBL01	57102	12.02				
VX0523WBS01	65596	12.02				
VX0523WBSD01	64515	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.