

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q2118 SAS No.: Q2118 SDG NO.: Q2118
 Lab File ID: VY022407.D Date Analyzed: 05/23/2025
 Instrument ID: MSVOA_Y Time Analyzed: 14:34
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	156128	13.352				
UPPER LIMIT	312256	13.852				
LOWER LIMIT	78064	12.852				
EPA SAMPLE NO.						
BP-VPB-182-GW-500-502	152683	13.35				
VY0523SBL01	167300	13.35				
VY0523SBS01	141917	13.35				
VY0523SBSD01	143885	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.