

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q2118 SAS No.: Q2118 SDG NO.: Q2118
 Lab File ID: VY022424.D Date Analyzed: 05/27/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:59
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	170568	13.346				
UPPER LIMIT	341136	13.846				
LOWER LIMIT	85284	12.846				
EPA SAMPLE NO.						
BP-VPB-182-GW-500-502RE	136693	13.35				
VY0527SBL01	193960	13.35				
VY0527SBS01	155764	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.