

ANALYTICAL RESULTS SUMMARY

SEMI-VOLATILE ORGANICS

PROJECT NAME : NWIRP BETHPAGE 112G08005-WE13

TETRA TECH NUS, INC.

661 Andersen Drive

Suite 200

Pittsburgh, PA - 15220-2745

Phone No: 412-921-7090

ORDER ID : Q2119

ATTENTION : Ernie Wu



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) SVOC-SIMGroup1- Case Narrative	4
3) Qualifier Page	6
4) QA Checklist	7
5) SVOC-SIMGroup1 Data	8
6) Shipping Document	37
6.1) CHAIN OF CUSTODY	38
6.2) Lab Certificate	39

1

2

3

4

5

6

Cover Page

Order ID : Q2119

Project ID : NWIRP Bethpage 112G08005-WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q2119-01
Q2119-02
Q2119-03

Client Sample Number

RW5-SP100-20250522
RW5-SP201-20250522
RW5-SP303-20250522

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 5/30/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager # Ernie Wu

Order ID # Q2119

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

3 Water samples were received on 05/22/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for RW5-SP201-20250522 [Terphenyl-d14 - 145%]. Failed surrogate is not associated with DOD, Therefore no further corrective action was taken.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

Samples RW5-SP100-20250522 was diluted was initially analyzed in sequence BN052725 where END CCAL Fail and also notices that sample required dilution therefore based on previous result lab analyzed sample directly with 2X. First analysis data provided as screening data in miscellaneous section and second analysis reported as final.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2119

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 05/30/2025

LAB CHRONICLE

OrderID:	Q2119	OrderDate:	5/22/2025 3:57:00 PM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	L41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2119-01	RW5-SP100-2025052 2	Water			05/22/25			05/22/25
			SVOC-SIMGroup1	8270-Modified		05/23/25	05/28/25	
Q2119-02	RW5-SP201-2025052 2	Water			05/22/25			05/22/25
			SVOC-SIMGroup1	8270-Modified		05/23/25	05/28/25	
Q2119-03	RW5-SP303-2025052 2	Water			05/22/25			05/22/25
			SVOC-SIMGroup1	8270-Modified		05/23/25	05/28/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: Q2119
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : RW5-SP100-20250522								
Q2119-01	RW5-SP100-20250522	WATER	1,4-Dioxane	5.600	0.13	0.4	0.4	ug/L
Total Svoc :				5.60				
Total Concentration:				5.60				



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/22/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	05/22/25
Client Sample ID:	RW5-SP100-20250522	SDG No.:	Q2119
Lab Sample ID:	Q2119-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037123.D	2	05/23/25 11:50	05/28/25 23:47	PB168155

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	5.60		0.13	0.40	0.40	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.27		30 - 150		68%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.33		30 - 150		82%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.26		55 - 111		65%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.27		53 - 106		67%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.52		58 - 132		131%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2490	7.611				
1146-65-2	Naphthalene-d8	7000	10.394				
15067-26-2	Acenaphthene-d10	3880	14.256				
1517-22-2	Phenanthrene-d10	7010	17.009				
1719-03-5	Chrysene-d12	4200	21.197				
1520-96-3	Perylene-d12	3570	23.395				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/22/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	05/22/25
Client Sample ID:	RW5-SP201-20250522	SDG No.:	Q2119
Lab Sample ID:	Q2119-02	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037104.D	1	05/23/25 11:50	05/28/25 11:53	PB168155

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.28		30 - 150		69%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.33		30 - 150		82%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.28		55 - 111		71%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.28		53 - 106		71%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.58	*	58 - 132		145%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1940	7.611				
1146-65-2	Naphthalene-d8	4980	10.394				
15067-26-2	Acenaphthene-d10	2650	14.256				
1517-22-2	Phenanthrene-d10	4790	17.009				
1719-03-5	Chrysene-d12	2910	21.198				
1520-96-3	Perylene-d12	2630	23.401				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/22/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	05/22/25
Client Sample ID:	RW5-SP303-20250522	SDG No.:	Q2119
Lab Sample ID:	Q2119-03	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037105.D	1	05/23/25 11:50	05/28/25 12:29	PB168155

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.28		30 - 150		70%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.38		30 - 150		96%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.30		55 - 111		74%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.28		53 - 106		71%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.53		58 - 132		132%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1930	7.611				
1146-65-2	Naphthalene-d8	4920	10.394				
15067-26-2	Acenaphthene-d10	2550	14.256				
1517-22-2	Phenanthrene-d10	4810	17.009				
1719-03-5	Chrysene-d12	3330	21.198				
1520-96-3	Perylene-d12	3070	23.398				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q2119

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168155BL	PB168155BL	2-Methylnaphthalene-d10	0.4	0.33	83		30	150
		Fluoranthene-d10	0.4	0.32	80		30	150
		Nitrobenzene-d5	0.4	0.32	80		55	111
		2-Fluorobiphenyl	0.4	0.35	86		53	106
		Terphenyl-d14	0.4	0.44	109		58	132
PB168155BS	PB168155BS	2-Methylnaphthalene-d10	0.4	0.38	95		30	150
		Fluoranthene-d10	0.4	0.30	75		30	150
		Nitrobenzene-d5	0.4	0.34	85		55	111
		2-Fluorobiphenyl	0.4	0.35	87		53	106
		Terphenyl-d14	0.4	0.42	105		58	132
PB168155BSD	PB168155BSD	2-Methylnaphthalene-d10	0.4	0.39	97		30	150
		Fluoranthene-d10	0.4	0.30	76		30	150
		Nitrobenzene-d5	0.4	0.36	89		55	111
		2-Fluorobiphenyl	0.4	0.37	93		53	106
		Terphenyl-d14	0.4	0.42	106		58	132
Q2119-01	RW5-SP100-20250522	2-Methylnaphthalene-d10	0.4	0.27	68		30	150
		Fluoranthene-d10	0.4	0.33	82		30	150
		Nitrobenzene-d5	0.4	0.26	65		55	111
		2-Fluorobiphenyl	0.4	0.27	67		53	106
		Terphenyl-d14	0.4	0.52	131		58	132
Q2119-02	RW5-SP201-20250522	2-Methylnaphthalene-d10	0.4	0.28	69		30	150
		Fluoranthene-d10	0.4	0.33	82		30	150
		Nitrobenzene-d5	0.4	0.28	71		55	111
		2-Fluorobiphenyl	0.4	0.28	71		53	106
		Terphenyl-d14	0.4	0.58	145	*	58	132
Q2119-03	RW5-SP303-20250522	2-Methylnaphthalene-d10	0.4	0.28	70		30	150
		Fluoranthene-d10	0.4	0.38	96		30	150
		Nitrobenzene-d5	0.4	0.30	74		55	111
		2-Fluorobiphenyl	0.4	0.28	71		53	106
		Terphenyl-d14	0.4	0.53	132		58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2119

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN037108.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168155BS	1,4-Dioxane	0.4	0.29	ug/L	73				70	130	

A
B
C
D
E
F
G

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2119

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN037109.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168155BSD	1,4-Dioxane	0.4	0.30	ug/L	75	3			70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168155BL

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q2119

SAS No.: Q2119 SDG NO.: Q2119

Lab File ID: BN037103.D

Lab Sample ID: PB168155BL

Instrument ID: BNA_N

Date Extracted: 05/23/2025

Matrix: (soil/water) Water

Date Analyzed: 05/28/2025

Level: (low/med) LOW

Time Analyzed: 11:17

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB168155BS	PB168155BS	BN037108.D	05/28/2025
RW5-SP201-20250522	Q2119-02	BN037104.D	05/28/2025
RW5-SP303-20250522	Q2119-03	BN037105.D	05/28/2025
PB168155BSD	PB168155BSD	BN037109.D	05/28/2025
RW5-SP100-20250522	Q2119-01	BN037123.D	05/28/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2119

SDG NO.: Q2119

Lab File ID: BN036998.D

DFTPP Injection Date: 05/13/2025

Instrument ID: BNA_N

DFTPP Injection Time: 17:02

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	62.8
68	Less than 2.0% of mass 69	0.8 (1.4) 1
69	Mass 69 relative abundance	55.6
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	52.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	8.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.4 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036999.D	05/13/2025	17:41
SSTDICC0.2	SSTDICC0.2	BN037000.D	05/13/2025	18:17
SSTDICCC0.4	SSTDICCC0.4	BN037001.D	05/13/2025	18:53
SSTDICC0.8	SSTDICC0.8	BN037002.D	05/13/2025	19:29
SSTDICC1.6	SSTDICC1.6	BN037003.D	05/13/2025	20:05
SSTDICC3.2	SSTDICC3.2	BN037004.D	05/13/2025	20:41
SSTDICC5.0	SSTDICC5.0	BN037005.D	05/13/2025	21:17

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2119

SDG NO.: Q2119

Lab File ID: BN037101.D

DFTPP Injection Date: 05/28/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:02

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	72.6
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	60.6
70	Less than 2.0% of mass 69	0.3 (0.5) 1
127	10.0 - 80.0% of mass 198	54.2
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	24.2
365	Greater than 1% of mass 198	4
441	Present, but less than mass 443	8.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10 (17.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037102.D	05/28/2025	10:41
PB168155BL	PB168155BL	BN037103.D	05/28/2025	11:17
RW5-SP201-20250522	Q2119-02	BN037104.D	05/28/2025	11:53
RW5-SP303-20250522	Q2119-03	BN037105.D	05/28/2025	12:29
PB168155BS	PB168155BS	BN037108.D	05/28/2025	14:17
PB168155BSD	PB168155BSD	BN037109.D	05/28/2025	14:53
SSTDCCC0.4EC	SSTDCCC0.4	BN037110.D	05/28/2025	15:54

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2119

SDG NO.: Q2119

Lab File ID: BN037111.D

DFTPP Injection Date: 05/28/2025

Instrument ID: BNA_N

DFTPP Injection Time: 16:31

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	75.5
68	Less than 2.0% of mass 69	0.9 (1.4) 1
69	Mass 69 relative abundance	62.5
70	Less than 2.0% of mass 69	0.4 (0.6) 1
127	10.0 - 80.0% of mass 198	56.1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	25.1
365	Greater than 1% of mass 198	4.6
441	Present, but less than mass 443	9.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	11.5 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037112.D	05/28/2025	17:11
RW5-SP100-20250522	Q2119-01	BN037123.D	05/28/2025	23:47
SSTDCCC0.4EC	SSTDCCC0.4	BN037126.D	05/29/2025	01:35

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2119 SAS No.: Q2119 SDG NO.: Q2119

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/28/2025

Lab File ID: BN037102.D Time Analyzed: 10:41

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2360	7.611	6377	10.38	3563	14.26
UPPER LIMIT	4720	8.111	12754	10.883	7126	14.756
LOWER LIMIT	1180	7.111	3188.5	9.883	1781.5	13.756
EPA SAMPLE NO.						
01 PB168155BS	1969	7.61	4930	10.39	2429	14.26
02 PB168155BSD	1893	7.61	4639	10.39	2241	14.26
03 PB168155BL	2460	7.61	5989	10.39	3067	14.26
04 RW5-SP201-20250522	1938	7.61	4981	10.39	2646	14.26
05 RW5-SP303-20250522	1933	7.61	4924	10.39	2554	14.26

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2119 SAS No.: Q2119 SDG NO.: Q2119

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/28/2025

Lab File ID: BN037102.D Time Analyzed: 10:41

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	6602	17.009	4312	21.197	3700	23.398
UPPER LIMIT	13204	17.509	8624	21.697	7400	23.898
LOWER LIMIT	3301	16.509	2156	20.697	1850	22.898
EPA SAMPLE NO.						
01 PB168155BS	4326	17.01	2746	21.20	2563	23.40
02 PB168155BSD	3936	17.01	2452	21.20	2400	23.40
03 PB168155BL	5493	17.01	3554	21.21	3269	23.40
04 RW5-SP201-20250522	4789	17.01	2905	21.20	2625	23.40
05 RW5-SP303-20250522	4806	17.01	3331	21.20	3065	23.40

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2119 SAS No.: Q2119 SDG NO.: Q2119

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/28/2025

Lab File ID: BN037112.D Time Analyzed: 17:11

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2022	7.611	5050	10.38	2480	14.26
UPPER LIMIT	4044	8.111	10100	10.883	4960	14.756
LOWER LIMIT	1011	7.111	2525	9.883	1240	13.756
EPA SAMPLE NO.						
01 RW5-SP100-20250522	2494	7.61	7003	10.39	3875	14.26

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2119 SAS No.: Q2119 SDG NO.: Q2119

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/28/2025

Lab File ID: BN037112.D Time Analyzed: 17:11

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	4231	17.009	3037	21.198	3091	23.395
UPPER LIMIT	8462	17.509	6074	21.698	6182	23.895
LOWER LIMIT	2115.5	16.509	1518.5	20.698	1545.5	22.895
EPA SAMPLE NO.						
01 RW5-SP100-20250522	7007	17.01	4199	21.20	3567	23.40

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:		
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:		
Client Sample ID:	PB168155BL		SDG No.:	Q2119	
Lab Sample ID:	PB168155BL		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037103.D	1	05/23/25 11:50	05/28/25 11:17	PB168155

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.33		30 - 150		83%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.32		30 - 150		80%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		80%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.35		53 - 106		86%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.44		58 - 132		109%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2460	7.611				
1146-65-2	Naphthalene-d8	5990	10.394				
15067-26-2	Acenaphthene-d10	3070	14.256				
1517-22-2	Phenanthrene-d10	5490	17.009				
1719-03-5	Chrysene-d12	3550	21.206				
1520-96-3	Perylene-d12	3270	23.404				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB168155BS		SDG No.:	Q2119
Lab Sample ID:	PB168155BS		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037108.D	1	05/23/25 11:50	05/28/25 14:17	PB168155

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.29		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.38		30 - 150		95%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.30		30 - 150		75%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		85%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.35		53 - 106		87%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.42		58 - 132		105%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1970	7.611				
1146-65-2	Naphthalene-d8	4930	10.394				
15067-26-2	Acenaphthene-d10	2430	14.256				
1517-22-2	Phenanthrene-d10	4330	17.009				
1719-03-5	Chrysene-d12	2750	21.198				
1520-96-3	Perylene-d12	2560	23.398				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB168155BSD		SDG No.:	Q2119
Lab Sample ID:	PB168155BSD		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037109.D	1	05/23/25 11:50	05/28/25 14:53	PB168155

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.30		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.39		30 - 150		97%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.30		30 - 150		76%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.36		55 - 111		89%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.37		53 - 106		93%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.42		58 - 132		106%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1890	7.611				
1146-65-2	Naphthalene-d8	4640	10.394				
15067-26-2	Acenaphthene-d10	2240	14.256				
1517-22-2	Phenanthrene-d10	3940	17.009				
1719-03-5	Chrysene-d12	2450	21.198				
1520-96-3	Perylene-d12	2400	23.398				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN051425.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 14 11:26:32 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN036999.D 0.2 =BN037000.D 0.4 =BN037001.D 0.8 =BN037002.D 1.6 =BN037003.D 3.2 =BN037004.D 5.0 =BN037005.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.510	0.512	0.487	0.514	0.467	0.454	0.491	5.25	
3)	n-Nitrosodimet...	1.465	0.974	0.980	0.971	1.075	0.967	0.950	1.054	17.59
4) S	2-Fluorophenol	1.101	1.134	1.024	1.093	0.964	0.971	1.048	6.87	
5) S	Phenol-d6	1.304	1.385	1.236	1.392	1.259	1.292	1.311	4.91	
6)	bis(2-Chloroet...	1.441	1.163	1.153	1.135	1.240	1.168	1.148	1.207	9.02
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.546	0.383	0.398	0.400	0.452	0.426	0.442	0.436	12.60
9)	Naphthalene	1.326	1.140	1.144	1.122	1.226	1.152	1.165	1.182	6.05
10)	Hexachlorobuta...	0.286	0.248	0.244	0.235	0.256	0.236	0.233	0.248	7.47
11) SURR2	Methylnaphth...	0.529	0.547	0.552	0.548	0.603	0.574	0.588	0.563	4.65
12)	2-Methylnaphth...	0.754	0.724	0.736	0.733	0.814	0.770	0.790	0.760	4.34
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.174	0.168	0.178	0.160	0.186	0.175	0.189	0.176	5.77
15) S	2-Fluorobiphenyl	1.912	1.801	1.901	1.802	1.927	1.672	1.807	1.832	4.90
16)	Acenaphthylene	1.906	1.838	1.894	1.849	2.071	1.997	2.075	1.947	5.14
17)	Acenaphthene	1.255	1.229	1.243	1.217	1.350	1.298	1.315	1.272	3.89
18)	Fluorene	1.602	1.581	1.635	1.611	1.779	1.721	1.752	1.669	4.80
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.060	0.073	0.079	0.102	0.103	0.124	0.090	26.02	
21)	4-Bromophenyl-...	0.243	0.246	0.250	0.247	0.262	0.261	0.259	0.253	3.13
22)	Hexachlorobenzene	0.267	0.269	0.281	0.259	0.281	0.270	0.267	0.270	3.03
23)	Atrazine	0.199	0.207	0.211	0.213	0.237	0.234	0.242	0.220	7.64
24)	Pentachlorophenol	0.133	0.134	0.141	0.137	0.159	0.162	0.177	0.149	11.45
25)	Phenanthrene	1.259	1.272	1.292	1.263	1.367	1.337	1.361	1.307	3.56
26)	Anthracene	1.099	1.104	1.166	1.130	1.269	1.259	1.300	1.190	7.13
27) SURR	Fluoranthene-d10	1.033	1.033	1.078	1.042	1.153	1.161	1.178	1.097	5.95
28)	Fluoranthene	1.461	1.439	1.500	1.496	1.670	1.672	1.693	1.562	7.13
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.744	1.708	1.727	1.656	1.790	1.641	1.711	1.711	2.96
31) S	Terphenyl-d14	0.897	0.844	0.871	0.822	0.891	0.816	0.848	0.856	3.73
32)	Benzo(a)anthra...	1.463	1.432	1.485	1.438	1.594	1.521	1.609	1.506	4.77
33)	Chrysene	1.655	1.559	1.616	1.532	1.653	1.560	1.576	1.593	3.05
34)	Bis(2-ethylhex...	0.955	0.919	0.906	0.855	0.941	0.903	1.011	0.927	5.27
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN051425.M

36)	Indeno(1,2,3-c...	1.511	1.613	1.645	1.568	1.687	1.732	1.680	1.634	4.65
37)	Benzo(b)fluora...	1.631	1.570	1.602	1.599	1.749	1.698	1.765	1.659	4.71
38)	Benzo(k)fluora...	1.539	1.538	1.642	1.601	1.770	1.661	1.719	1.639	5.34
39) C	Benzo(a)pyrene	1.380	1.343	1.381	1.331	1.486	1.444	1.486	1.407	4.59
40)	Dibenzo(a,h)an...	1.116	1.232	1.273	1.237	1.340	1.376	1.334	1.272	6.90
41)	Benzo(g,h,i)pe...	1.299	1.407	1.424	1.330	1.403	1.439	1.376	1.383	3.72

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2119 SAS No.: Q2119 SDG No.: Q2119

Instrument ID: BNA_N Calibration Date/Time: 05/28/2025 10:41

Lab File ID: BN037102.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 17:41 21:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.563	0.578		2.7	20.0
Fluoranthene-d10	1.097	0.983		-10.4	20.0
2-Fluorophenol	1.048	0.917		-12.5	20.0
Phenol-d6	1.311	1.136		-13.3	20.0
Nitrobenzene-d5	0.436	0.404		-7.3	20.0
2-Fluorobiphenyl	1.832	1.781		-2.8	20.0
2,4,6-Tribromophenol	0.176	0.150		-14.8	20.0
Terphenyl-d14	0.856	0.973		13.7	20.0
1,4-Dioxane	0.491	0.461		-6.1	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2119 SAS No.: Q2119 SDG No.: Q2119

Instrument ID: BNA_N Calibration Date/Time: 05/28/2025 15:54

Lab File ID: BN037110.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 17:41 21:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.563	0.578		2.7	50.0
Fluoranthene-d10	1.097	1.001		-8.8	50.0
2-Fluorophenol	1.048	0.942		-10.1	50.0
Phenol-d6	1.311	1.119		-14.6	50.0
Nitrobenzene-d5	0.436	0.411		-5.7	50.0
2-Fluorobiphenyl	1.832	1.776		-3.1	50.0
2,4,6-Tribromophenol	0.176	0.163		-7.4	50.0
Terphenyl-d14	0.856	0.988		15.4	50.0
1,4-Dioxane	0.491	0.478		-2.6	50.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2119 SAS No.: Q2119 SDG No.: Q2119

Instrument ID: BNA_N Calibration Date/Time: 05/28/2025 17:11

Lab File ID: BN037112.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 17:41 21:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.563	0.552		-2.0	20.0
Fluoranthene-d10	1.097	0.986		-10.1	20.0
2-Fluorophenol	1.048	0.938		-10.5	20.0
Phenol-d6	1.311	1.123		-14.3	20.0
Nitrobenzene-d5	0.436	0.426		-2.3	20.0
2-Fluorobiphenyl	1.832	1.889		3.1	20.0
2,4,6-Tribromophenol	0.176	0.145		-17.6	20.0
Terphenyl-d14	0.856	0.888		3.7	20.0
1,4-Dioxane	0.491	0.489		-0.4	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2119 SAS No.: Q2119 SDG No.: Q2119

Instrument ID: BNA_N Calibration Date/Time: 05/29/2025 01:35

Lab File ID: BN037126.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 17:41 21:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.563	0.577		2.5	50.0
Fluoranthene-d10	1.097	1.019		-7.1	50.0
2-Fluorophenol	1.048	0.946		-9.7	50.0
Phenol-d6	1.311	1.170		-10.8	50.0
Nitrobenzene-d5	0.436	0.407		-6.7	50.0
2-Fluorobiphenyl	1.832	1.641		-10.4	50.0
2,4,6-Tribromophenol	0.176	0.153		-13.1	50.0
Terphenyl-d14	0.856	1.004		17.3	50.0
1,4-Dioxane	0.491	0.459		-6.5	50.0

All other compounds must meet a minimum RRF of 0.010.



SHIPPING DOCUMENTS

38 of 39

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488