

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052825\
 Data File : BN037126.D
 Acq On : 29 May 2025 01:35
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 29 03:44:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 14 11:26:32 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	117	0.00
2	1,4-Dioxane	0.400	0.374	6.5	105	0.00
3	n-Nitrosodimethylamine	0.400	0.403	-0.8	127	0.00
4 S	2-Fluorophenol	0.400	0.361	9.8	98	-0.01
5 S	Phenol-d6	0.400	0.357	10.8	99	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.377	5.8	116	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	121	-0.02
8 S	Nitrobenzene-d5	0.400	0.374	6.5	124	-0.01
9	Naphthalene	0.400	0.377	5.8	118	-0.01
10	Hexachlorobutadiene	0.400	0.408	-2.0	126	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.410	-2.5	127	0.00
12	2-Methylnaphthalene	0.400	0.374	6.5	117	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	125	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.349	12.8	108	-0.01
15 S	2-Fluorobiphenyl	0.400	0.358	10.5	108	-0.02
16	Acenaphthylene	0.400	0.365	8.8	117	-0.02
17	Acenaphthene	0.400	0.369	7.8	118	-0.01
18	Fluorene	0.400	0.367	8.3	117	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	117	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.410	-2.5	127	-0.01
21	4-Bromophenyl-phenylether	0.400	0.398	0.5	118	-0.01
22	Hexachlorobenzene	0.400	0.423	-5.7	119	-0.01
23	Atrazine	0.400	0.375	6.3	115	-0.01
24	Pentachlorophenol	0.400	0.342	14.5	106	-0.01
25	Phenanthrene	0.400	0.379	5.3	112	-0.01
26	Anthracene	0.400	0.364	9.0	109	-0.01
27 SURR	Fluoranthene-d10	0.400	0.372	7.0	111	0.00
28	Fluoranthene	0.400	0.339	15.3	103	-0.01
29 I	Chrysene-d12	0.400	0.400	0.0	88	0.00
30	Pyrene	0.400	0.460	-15.0	101	-0.01
31 S	Terphenyl-d14	0.400	0.469	-17.2	102	-0.01
32	Benzo(a)anthracene	0.400	0.365	8.8	82	-0.02
33	Chrysene	0.400	0.394	1.5	86	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.401	-0.3	91	0.00
35 I	Perylene-d12	0.400	0.400	0.0	76	-0.02
36	Indeno(1,2,3-cd)pyrene	0.400	0.381	4.8	72	-0.03
37	Benzo(b)fluoranthene	0.400	0.371	7.3	73	-0.02
38	Benzo(k)fluoranthene	0.400	0.394	1.5	74	-0.02
39 C	Benzo(a)pyrene	0.400	0.370	7.5	71	-0.02
40	Dibenzo(a,h)anthracene	0.400	0.375	6.3	71	-0.03
41	Benzo(g,h,i)perylene	0.400	0.402	-0.5	74	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4