

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2119 SAS No.: Q2119 SDG NO.: Q2119

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/28/2025

Lab File ID: BN037112.D Time Analyzed: 17:11

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2022	7.611	5050	10.38	2480	14.26
UPPER LIMIT	4044	8.111	10100	10.883	4960	14.756
LOWER LIMIT	1011	7.111	2525	9.883	1240	13.756
EPA SAMPLE NO.						
01 RW5-SP100-20250522	2494	7.61	7003	10.39	3875	14.26

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.