

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2119

SDG NO.: Q2119

Lab File ID: BN036998.D

DFTPP Injection Date: 05/13/2025

Instrument ID: BNA_N

DFTPP Injection Time: 17:02

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	62.8
68	Less than 2.0% of mass 69	0.8 (1.4) 1
69	Mass 69 relative abundance	55.6
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	52.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	8.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.4 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036999.D	05/13/2025	17:41
SSTDICC0.2	SSTDICC0.2	BN037000.D	05/13/2025	18:17
SSTDICCC0.4	SSTDICCC0.4	BN037001.D	05/13/2025	18:53
SSTDICC0.8	SSTDICC0.8	BN037002.D	05/13/2025	19:29
SSTDICC1.6	SSTDICC1.6	BN037003.D	05/13/2025	20:05
SSTDICC3.2	SSTDICC3.2	BN037004.D	05/13/2025	20:41
SSTDICC5.0	SSTDICC5.0	BN037005.D	05/13/2025	21:17