ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

MATRIX: Water

ORDER ID: Q2120

METH	OD: 8270-Modified/3510			
1. 2.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks) GC/MS Tuning Specifications. DFTPP Meet Criteria.	NA	NO	YES ✓
	(NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			√
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.			✓
	The Initial Calibration met the requirements. The Continuous Calibration met the requirements.			
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria		\checkmark	
	If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
	The Surrogate recoveries met the acceptable criteria except for VPB182-HYD-20250522 [2-Methylnaphthalene-d10 - 6%, Fluoranthene-d10 - 23%, Terphenyl-d14 - 140%], due to matrix interference which can be observed by the abnormal chromatogram, and due to the limited volume of this samples they can not be re-			
	extracted. and BP-VPB-182-EB-20250522 [Terphenyl-d14 - 161%]. Failed surrogate is not associated with DOD, Therefor no further corrective action was taken.			

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

QA RE	VIEW Date	_
the Initi	ial Calibration curve for SW-846 analysis.	
	culated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20	% for
	value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount ac	
Please ı	use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the	ne
The not	t QT review data is reported in the Miscellaneous.	
Summa	oratory certifies that the all-electronic diskette deliverable exactly match the data ry forms (i.e. Form Is)." rm 6 is not included in the data package because the Initial Calibration was performed using 7 points.	
ADDIT	TONAL COMMENTS:	
	If not met, list number of days exceeded for each sample:	
11.	Analysis Holding Time Met	✓
	If not met, list number of days exceeded for each sample:	
10.	Extraction Holding Time Met	✓
	Comments:	
9.	Internal Standard Area/Retention Time Shift Meet Criteria	✓
	The Blank Spike met requirements for all samples. The Blank Spike Duplicate met requirements for all samples.	
	If not met, list those compounds and their recoveries which fall outside the acceptable range.	
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria	✓