

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2120

SDG NO.: Q2120

Lab File ID: BN037101.D

DFTPP Injection Date: 05/28/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:02

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	72.6
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	60.6
70	Less than 2.0% of mass 69	0.3 (0.5) 1
127	10.0 - 80.0% of mass 198	54.2
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	24.2
365	Greater than 1% of mass 198	4
441	Present, but less than mass 443	8.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10 (17.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037102.D	05/28/2025	10:41
PB168155BL	PB168155BL	BN037103.D	05/28/2025	11:17
VPB182-HYD-20250522	Q2120-01	BN037106.D	05/28/2025	13:05
BP-VPB-182-EB-20250522	Q2120-02	BN037107.D	05/28/2025	13:41
PB168155BS	PB168155BS	BN037108.D	05/28/2025	14:17
PB168155BSD	PB168155BSD	BN037109.D	05/28/2025	14:53
SSTDCCC0.4EC	SSTDCCC0.4	BN037110.D	05/28/2025	15:54