

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2120 SAS No.: Q2120 SDG No.: Q2120

Instrument ID: MSVOA_X Calibration Date/Time: 05/23/2025 19:55

Lab File ID: VX046357.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.742	0.828	0.1	11.59	50
Vinyl Chloride	0.691	0.742		7.38	50
Bromomethane	0.320	0.337		5.31	50
Chloroethane	0.369	0.428		15.99	50
Trichlorofluoromethane	1.021	1.173		14.89	50
1,1,2-Trichlorotrifluoroethane	0.632	0.715		13.13	50
1,1-Dichloroethene	0.593	0.657		10.79	50
Acetone	0.374	0.428		14.44	50
Carbon Disulfide	1.406	1.429		1.64	50
Methyl tert-butyl Ether	2.079	2.383		14.62	50
Methylene Chloride	0.716	0.774		8.1	50
trans-1,2-Dichloroethene	0.596	0.665		11.58	50
1,1-Dichloroethane	1.219	1.413	0.1	15.91	50
2-Butanone	0.543	0.643		18.42	50
Carbon Tetrachloride	0.544	0.614		12.87	50
cis-1,2-Dichloroethene	0.718	0.826		15.04	50
Chloroform	1.271	1.489		17.15	50
1,1,1-Trichloroethane	1.101	1.253		13.81	50
Methylcyclohexane	0.623	0.674		8.19	50
Benzene	1.417	1.650		16.44	50
1,2-Dichloroethane	0.612	0.717		17.16	50
Trichloroethene	0.341	0.379		11.14	50
1,2-Dichloropropane	0.352	0.422		19.89	50
Bromodichloromethane	0.547	0.652		19.2	50
4-Methyl-2-Pentanone	0.605	0.739		22.15	50
Toluene	0.869	1.003		15.42	50
t-1,3-Dichloropropene	0.487	0.559		14.78	50
cis-1,3-Dichloropropene	0.538	0.614		14.13	50
1,1,2-Trichloroethane	0.343	0.411		19.83	50
2-Hexanone	0.448	0.551		22.99	50
Dibromochloromethane	0.376	0.451		19.95	50
Tetrachloroethene	0.354	0.391		10.45	50
Chlorobenzene	1.094	1.215	0.3	11.06	50
Ethyl Benzene	1.929	2.216		14.88	50
m/p-Xylenes	0.706	0.808		14.45	50
o-Xylene	0.688	0.787		14.39	50
Styrene	1.127	1.347		19.52	50
Bromoform	0.281	0.310	0.1	10.32	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2120 SAS No.: Q2120 SDG No.: Q2120

Instrument ID: MSVOA_X Calibration Date/Time: 05/23/2025 19:55

Lab File ID: VX046357.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.893	4.443		14.13	50
1,1,2,2-Tetrachloroethane	1.364	1.471	0.3	7.84	50
1,3-Dichlorobenzene	1.649	1.817		10.19	50
1,4-Dichlorobenzene	1.684	1.802		7.01	50
1,2-Dichlorobenzene	1.655	1.854		12.02	50
1,2-Dichloroethane-d4	0.932	0.992		6.44	50
Dibromofluoromethane	0.360	0.392		8.89	50
Toluene-d8	1.246	1.306		4.82	50
4-Bromofluorobenzene	0.478	0.512		7.11	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.