

Surrogate Summary

SDG No.: Q2120

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2120-01	VPB182-HYD-20250522	1,2-Dichloroethane-d4	50	54.0	108	81	118
		Dibromofluoromethane	50	50.5	101	80	119
		Toluene-d8	50	49.7	99	89	112
		4-Bromofluorobenzene	50	52.5	105	85	114
Q2120-02	BP-VPB-182-EB-20250522	1,2-Dichloroethane-d4	50	53.6	107	81	118
		Dibromofluoromethane	50	50.5	101	80	119
		Toluene-d8	50	50.7	101	89	112
		4-Bromofluorobenzene	50	51.0	102	85	114
VX0523WBL01	VX0523WBL01	1,2-Dichloroethane-d4	50	53.2	106	81	118
		Dibromofluoromethane	50	52.0	104	80	119
		Toluene-d8	50	51.1	102	89	112
		4-Bromofluorobenzene	50	51.9	104	85	114
VX0523WBS01	VX0523WBS01	1,2-Dichloroethane-d4	50	50.5	101	81	118
		Dibromofluoromethane	50	52.4	105	80	119
		Toluene-d8	50	50.7	101	89	112
		4-Bromofluorobenzene	50	51.4	103	85	114
VX0523WBSD0	VX0523WBSD01	1,2-Dichloroethane-d4	50	50.2	100	81	118
		Dibromofluoromethane	50	51.1	102	80	119
		Toluene-d8	50	50.3	101	89	112
		4-Bromofluorobenzene	50	51.8	104	85	114