

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2120 SAS No.: Q2120 SDG NO.: Q2120
Lab File ID: VX046331.D Date Analyzed: 05/23/2025
Instrument ID: MSVOA_X Time Analyzed: 09:24
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	89943	5.54	154938	6.76	137123	10.05
UPPER LIMIT	179886	6.044	309876	7.257	274246	10.549
LOWER LIMIT	44971.5	5.044	77469	6.257	68561.5	9.549
EPA SAMPLE NO.						
VPB182-HYD-20250522	61857	5.55	124865	6.76	120584	10.05
BP-VPB-182-EB-20250522	63849	5.55	127611	6.76	124029	10.05
VX0523WBL01	69169	5.55	136383	6.76	129888	10.05
VX0523WBS01	88801	5.54	154286	6.75	135355	10.05
VX0523WBSD01	88781	5.55	156018	6.76	138561	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.