

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2120 SAS No.: Q2120 SDG NO.: Q2120
Lab File ID: VX046331.D Date Analyzed: 05/23/2025
Instrument ID: MSVOA_X Time Analyzed: 09:24
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	66629	12.018				
UPPER LIMIT	133258	12.518				
LOWER LIMIT	33314.5	11.518				
EPA SAMPLE NO.						
VPB182-HYD-20250522	52108	12.02				
BP-VPB-182-EB-20250522	52711	12.02				
VX0523WBL01	57102	12.02				
VX0523WBS01	65596	12.02				
VX0523WBSD01	64515	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.