

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q2123 SAS No.: Q2123 SDG NO.: Q2123
 Lab File ID: VX046331.D Date Analyzed: 05/23/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:24
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	66629	12.018				
UPPER LIMIT	133258	12.518				
LOWER LIMIT	33314.5	11.518				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	43948	12.02				
STORAGE-BLANK-WATER-REF-5	58155	12.02				
STORAGE-BLANK-WATER-REF-4	44717	12.02				
STORAGE-BLANK-SAMPLE-MANAGEMENT	45680	12.02				
VX0523WBL01	57102	12.02				
VX0523WBS01	65596	12.02				
VX0523WBSD01	64515	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.