

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052925\
 Data File : VL042588.D
 Acq On : 29 May 2025 16:50
 Operator : SY/MD
 Sample : VL0529ABL02
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0529ABL02

Quant Time: May 30 02:20:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 30 02:01:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.797	49	102651	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.971	114	295997	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.894	117	264969	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.386	95	199114	10.025	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.200%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052925\
Data File : VL042588.D
Acq On : 29 May 2025 16:50
Operator : SY/MD
Sample : VL0529ABL02
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0529ABL02

Quant Time: May 30 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri May 30 02:01:56 2025
Response via : Initial Calibration









#1
 Bromochloromethane
 Concen: 10.000 ppbv
 RT: 2.797 min Scan# 81
 Delta R.T. 0.006 min
 Lab File: VL042588.D
 Acq: 29 May 2025 16:50




Instrument :
 MSVOA_L
 ClientSampleId :
 VL0529ABL02


Tgt Ion: 49 Resp: 102651
 Ion Ratio Lower Upper
 49 100
 128 45.2 24.2 72.6
 130 58.5 30.4 91.2



#33
 1,4-Difluorobenzene
 Concen: 10.000 ppbv
 RT: 3.971 min Scan# 1200
 Delta R.T. 0.010 min
 Lab File: VL042588.D
 Acq: 29 May 2025 16:50

Tgt Ion: 114 Resp: 295997
 Ion Ratio Lower Upper
 114 100
 63 23.9 19.0 28.6
 88 17.6 13.6 20.4









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.894 min Scan# 21
Delta R.T. 0.006 min
Lab File: VL042588.D
Acq: 29 May 2025 16:50

Instrument :
MSVOA_1
ClientSampleId :
VL0529ABL02

Tgt Ion:117 Resp: 264969
Ion Ratio Lower Upper
117 100
82 62.4 51.2 76.8
119 32.4 26.3 39.5



#68
1-Bromo-4-Fluorobenzene
Concen: 10.025 ppbv
RT: 10.386 min Scan# 3182
Delta R.T. 0.006 min
Lab File: VL042588.D
Acq: 29 May 2025 16:50

Tgt Ion: 95 Resp: 199114
Ion Ratio Lower Upper
95 100
174 66.7 53.1 79.7
175 5.1 4.2 6.2

