

Continuing Calibration Report for SequenceID : FC052825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069029.D

Aliphatic C9-C12	6313437.000	60.000	3.339	6.638	105223.950	105717.491	0.467
Aliphatic C12-C16	4039143.000	40.000	6.639	10.038	100978.575	101985.435	0.987
Aliphatic C16-C21	5809366.000	60.000	10.039	13.404	96822.767	97865.081	1.065
Aliphatic C21-C28	7842342.000	80.000	13.405	17.066	98029.275	93803.915	-4.504
Aliphatic C28-C40	11834102.000	120.000	17.067	22.060	98617.517	94439.035	-4.425
Aliphatic EPH	35838390.000	360.000	3.339	22.060	99551.083	97587.136	-2.013