

Continuing Calibration Report for SequenceID : FC052825AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FC069041.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 6298433.000  | 60.000  | 3.339  | 6.638  | 104973.883 | 105717.491 | 0.703  |
| Aliphatic C12-C16 | 3921712.000  | 40.000  | 6.639  | 10.038 | 98042.800  | 101985.435 | 3.866  |
| Aliphatic C16-C21 | 5539983.000  | 60.000  | 10.039 | 13.404 | 92333.050  | 97865.081  | 5.653  |
| Aliphatic C21-C28 | 7504980.000  | 80.000  | 13.405 | 17.066 | 93812.250  | 93803.915  | -0.009 |
| Aliphatic C28-C40 | 11206658.000 | 120.000 | 17.067 | 22.060 | 93388.817  | 94439.035  | 1.112  |
| Aliphatic EPH     | 34471766.000 | 360.000 | 3.339  | 22.060 | 95754.906  | 97587.136  | 1.878  |