

Continuing Calibration Report for SequenceID : FC052925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069044.D

Aliphatic C9-C12	6533923.000	60.000	3.340	6.639	108898.717	105717.491	-3.009
Aliphatic C12-C16	4085377.000	40.000	6.640	10.038	102134.425	101985.435	-0.146
Aliphatic C16-C21	5764892.000	60.000	10.039	13.405	96081.533	97865.081	1.822
Aliphatic C21-C28	7723703.000	80.000	13.406	17.069	96546.288	93803.915	-2.924
Aliphatic C28-C40	11446094.000	120.000	17.070	22.063	95384.117	94439.035	-1.001
Aliphatic EPH	35553989.000	360.000	3.340	22.063	98761.081	97587.136	-1.203