

Continuing Calibration Report for SequenceID : FC052925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069049.D

Aliphatic C9-C12	6414070.000	60.000	3.340	6.639	106901.167	105717.491	-1.120
Aliphatic C12-C16	4116698.000	40.000	6.640	10.038	102917.450	101985.435	-0.914
Aliphatic C16-C21	5871755.000	60.000	10.039	13.405	97862.583	97865.081	0.003
Aliphatic C21-C28	7782994.000	80.000	13.406	17.069	97287.425	93803.915	-3.714
Aliphatic C28-C40	11520103.000	120.000	17.070	22.063	96000.858	94439.035	-1.654
Aliphatic EPH	35705620.000	360.000	3.340	22.063	99182.278	97587.136	-1.635