

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2125 SAS No.: Q2125 SDG No.: Q2125
Instrument ID: BNA_M Calibration Date(s): 06/05/2025 06/05/2025
Calibration Time(s): 09:20 13:56

LAB FILE ID:		RRF2.5 = BM050194.D		RRF005 = BM050195.D		RRF010 = BM050196.D		
		RRF020 = BM050197.D		RRF040 = BM050198.D		RRF050 = BM050199.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.251	1.193	1.192	1.145	1.238	1.199	3.5
Phenol-d6		1.569	1.521	1.617	1.565	1.699	1.578	4.0
Nitrobenzene-d5		0.359	0.362	0.392	0.377	0.414	0.385	5.4
Naphthalene		1.111	1.041	1.056	0.977	1.061	1.034	4.8
2-Methylnaphthalene		0.625	0.592	0.643	0.615	0.679	0.624	4.8
2-Fluorobiphenyl		1.589	1.536	1.532	1.377	1.463	1.477	5.7
2,4,6-Tribromophenol		0.210	0.213	0.238	0.231	0.254	0.227	7.0
Terphenyl-d14		1.148	1.088	1.165	1.010	1.064	1.055	8.1