

SOP ID : M1312-SPLP-10
SDG No : N/A
Weigh By : JP
Balance ID : WC SC-7
pH Meter ID : WC PH METER-1
Extraction By : JP
Filter By : JP
Pipette ID : WC
Tumbler ID : T-2
TCLP Filter ID : 115525

Start Prep Date : 06/04/2025 Time : 15:00
End Prep Date : 06/05/2025 Time : 09:20
Combination Ratio : 20
ZHE Cleaning Batch : ⁷⁰N/A
Initial Room Temperature: 24 °C
Final Room Temperature: 22 °C
TCLP Technician Signature : JP
Supervisor By : 12

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
SPLP FLUID	WP112802	N/A
N/A	N/A	N/A
HNO3-TCLP, 1N	N/A	WP112799
pH Strips	N/A	W1931, W1934, W3171, W3172
N/A	N/A	N/A
1 Liter Amber	N/A	90924-08
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

TUMBLER T-2 checked, 30 rpm.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
06/05/25 11:30	<u>JP</u> <u>100 AOCm</u>	<u>RS</u> <u>1E+1</u>
	Preparation Group	Analysis Group <u>DEP</u>

SPLP EXTRACTION LOGPAGE

PB168273

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB168273TB	LEB273	17	N/A	2000	N/A	N/A	N/A	4.23	N/A	T-2
Q2125-08	GSB3	16	100.05	2000	N/A	N/A	N/A	5.0	N/A	T-2

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB168273TB	LEB273	N/A	N/A	N/A	N/A	N/A	N/A
Q2125-08	GSB3	N/A	N/A	N/A	N/A	100	N/A

Hot Block ID : N/A

Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
PB168273TB	LEB273	N/A	N/A	N/A	N/A	#1	4.23
Q2125-08	GSB3	N/A	N/A	N/A	N/A	#1	4.23