

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2125 SAS No.: Q2125 SDG NO.: Q2125

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/09/2025

Lab File ID: BM050236.D Time Analyzed: 10:47

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	371799	7.769	1449700	10.56	826607	14.40
UPPER LIMIT	743598	8.269	2899400	11.057	1653210	14.898
LOWER LIMIT	185900	7.269	724850	10.057	413304	13.898
EPA SAMPLE NO.						
01 PB168340BL	308898	7.77	1157440	10.56	647360	14.40
02 PB168340BS	320774	7.77	1257050	10.56	743659	14.40
03 PB168340TB	295930	7.77	1103530	10.55	626432	14.40
04 GSB3	258819	7.77	1009930	10.55	581596	14.40
05 GSB3MS	289351	7.77	1143580	10.56	672610	14.40
06 GSB3MSD	281877	7.77	1107270	10.56	644595	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.