

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2125 SAS No.: Q2125 SDG No.: Q2125
Instrument ID: BNA_P Calibration Date(s): 06/06/2025 06/06/2025
Calibration Time(s): 10:30 15:18

LAB FILE ID:		RRF2.5 = BP024860.D			RRF005 = BP024861.D		RRF010 = BP024862.D	
		RRF020 = BP024863.D			RRF040 = BP024864.D		RRF050 = BP024865.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.127	1.139	1.207	1.163	1.283	1.198	4.9
Phenol-d6		1.507	1.528	1.588	1.545	1.676	1.585	4.5
Nitrobenzene-d5		0.407	0.397	0.423	0.404	0.444	0.412	4.9
Naphthalene		1.071	1.022	1.044	0.989	1.079	1.025	4.9
2-Methylnaphthalene		0.659	0.638	0.659	0.633	0.696	0.650	4.5
2-Fluorobiphenyl		1.542	1.507	1.517	1.390	1.563	1.485	4.4
2,4,6-Tribromophenol		0.256	0.264	0.279	0.267	0.298	0.277	5.3
Terphenyl-d14		1.178	1.073	1.146	1.089	1.164	1.116	4.6